

Research Use Only. Not for use in diagnostic procedures.

## Glutathione Donor Beads

**Product No.:** 6765300, 6765301, 6765302, 6765303

**Lot No.:** 2797940

### Material Provided

<b>Format:</b>	6765300	1 mg*
	6765301	5 mg
	6765302	25 mg
	6765303	XXX mg (LQC)

\*Note: 1mg translates into 2,000 assay points based on a final bead concentration of 20 µg/mL in a 25 µL/well reaction volume.

**Manufacturing Date:** November 20, 2020

#### Kit Components:

Component	<b>6765300 ( 1 mg )</b>	<b>6765301 ( 5 mg )</b>	<b>6765302 ( 25 mg )</b>	<b>6765303 ( LQC mg )</b>
<b>Glutathione Donor Beads at 5 mg/mL in 25 mM Hepes, 0.1% Tween-20, 0.05% Kathon, pH 7.4</b>	1 x 200 µL (6765300)	1 x 1 mL (6765301)	1 x 5 mL (6765302)	___ x ___ mL (6765303)

### Product Information

**Intended use:** AlphaScreen® Glutathione Donor beads specifically bind to GST-tagged proteins. This product should be used in conjunction with an AlphaScreen® or AlphaLISA® Acceptor bead.

**Stability:** This product is stable for at least **24 months** from the manufacturing date if used and stored under recommended conditions.

**Storage Conditions:** Store undiluted at 4°C protected from light. Freeze-thaw is not recommended and can cause the beads to form aggregates.

**Recommended use:** AlphaScreen® Donor beads are light-sensitive. All Alpha assays using the Donor beads should be performed under subdued laboratory lighting (< 100 lux). Green filters (LEE 090 filters (preferred) or Roscolux filters #389 from Rosco) can be applied to light fixtures.

For additional information on running AlphaScreen® assays or on potential interfering compounds, please visit our website: [www.perkinelmer.com/AlphaTech](http://www.perkinelmer.com/AlphaTech)

## Quality Control

Maximum signal and background counts were determined in an AlphaScreen® assay using anti-GST acceptor beads from kit 6760603 with 100 µM GST. Assay was detected on an EnVision® HTS Alpha instrument. Current lot was assayed in parallel with a previous (Reference) Glutathione donor beads lot to ensure consistency. A Ratio to Reference (RTR) was calculated from these results. We certify that these results meet our requirements.

<u>TEST GST</u>	<u>RESULTS</u>	<u>REFERENCE</u>	<u>RTR</u>
Maximum signal	40456	34530	1.17
Background	980	710	

## Suggested Materials and Instrumentation

Please visit our website

[www.perkinelmer.com/AlphaTech](http://www.perkinelmer.com/AlphaTech)

**This product is not for resale or distribution except by authorized distributors.**

**LIMITED WARRANTY:** PerkinElmer BioSignal Inc. warrants that, at the time of shipment, the products sold by it are free from defects in material and workmanship and conform to specifications which accompany the product. PerkinElmer BioSignal Inc. makes no other warranty, express or implied with respect to the products, including any warranty of merchantability or fitness for any particular purpose. Notification of any breach of warranty must be made within 60 days of receipt unless otherwise provided in writing by PerkinElmer BioSignal Inc. No claim shall be honored if the customer fails to notify PerkinElmer BioSignal Inc. within the period specified. The sole and exclusive remedy of the customer for any liability of PerkinElmer BioSignal Inc. of any kind including liability based upon warranty (express or implied whether contained herein or elsewhere), strict liability contract or otherwise is limited to the replacement of the goods or the refunds of the invoice price of goods. PerkinElmer BioSignal Inc. shall not in any case be liable for special, incidental or consequential damages of any kind.

PerkinElmer, Inc.  
940 Winter Street  
Waltham, MA 02451 USA  
P: (800) 762-4000 or  
(+1) 203-925-4602  
[www.perkinelmer.com](http://www.perkinelmer.com)



For a complete listing of our global offices, visit [www.perkinelmer.com/ContactUs](http://www.perkinelmer.com/ContactUs)

Copyright ©2009, PerkinElmer, Inc. All rights reserved. PerkinElmer® is a registered trademark of PerkinElmer, Inc. All other trademarks are the property of their respective owners.