

Research Use Only. Not for use in diagnostic procedures.

AlphaLISA 5X Epigenetics Buffer 1 Kit

Product No.: AL008C (10 mL Kit)
AL008F (100 mL Kit)

Lot No.: 2914329

Material Provided

Kit Components	10 mL Kit	100 mL Kit
AlphaLISA 5X Epigenetics Buffer 1	AL008C1 (10 mL)	AL008F1 (100 mL)
AlphaLISA 30X Epigenetics Buffer Supplement	AL008C2 (1.7 mL)	AL008F2 (17 mL)

Manufacturing Date: August 31, 2021

Product Information

Description: This product is a two-component buffer kit specifically formulated for the detection of histone H3-derived peptides in AlphaLISA Epigenetics assays.

Application: Once prepared, the 1X AlphaLISA Epigenetics Buffer 1 is used for diluting AlphaLISA Acceptor and Alpha Donor beads. It prevents the non-specific binding of highly-charged histone peptides to the Acceptor and Alpha Donor beads in AlphaLISA Epigenetics assays.

Storage: Store kit components at 4°C.

Stability: The undiluted stock solutions are stable for at least 12 months from the manufacturing date when stored in their original packaging at the recommended storage conditions. **Upon preparation, the 1X AlphaLISA Epigenetics Buffer 1 should be used within 16 hours.**

Preparation: To 7.67 volumes of ultrapure water (18 MΩ·cm), add 2 volumes of 5X Epigenetics Buffer 1 and 0.33 volume of 30X Buffer Supplement. **The 1X AlphaLISA Epigenetics Buffer 1 may appear slightly cloudy or precipitation may occur. However, this will not affect assay performance. DO NOT heat above 30°C as this could significantly affect performance.**

Please visit our website at www.perkinelmer.com/epigenetics.

This product is not for resale or distribution except by authorized distributors.

LIMITED WARRANTY: PerkinElmer BioSignal Inc. warrants that, at the time of shipment, the products sold by it are free from defects in material and workmanship and conform to specifications which accompany the product. PerkinElmer BioSignal Inc. makes no other warranty, express or implied with respect to the products, including any warranty of merchantability or fitness for any particular purpose. Notification of any breach of warranty must be made within 60 days of receipt unless otherwise provided in writing by PerkinElmer BioSignal Inc. No claim shall be honored if the customer fails to notify PerkinElmer BioSignal Inc. within the period specified. The sole and exclusive remedy of the customer for any liability of PerkinElmer BioSignal Inc. of any kind including liability based upon warranty (express or implied whether contained herein or elsewhere), strict liability contract or otherwise is limited to the replacement of the goods or the refunds of the invoice price of goods. PerkinElmer BioSignal Inc. shall not in any case be liable for special, incidental or consequential damages of any kind.

PerkinElmer, Inc.
940 Winter Street
Waltham, MA 02451 USA
P: (800) 762-4000 or
(+1) 203-925-4602
www.perkinelmer.com

