

Certificate of Analysis

Caution: Research use only.

AlphaLISA® SureFire® Ultra™ Assay Kit p70S6K Total

PRODUCT NUMBER:	ALSU-TP70-A 500
LOT NUMBER:	U7601
MANUFACTURING DATE:	February 10, 2021

MATERIAL PROVIDED

CONTENTS: Reaction Buffer 1 -Ultra, Reaction Buffer 2 -Ultra, Activation Buffer -Ultra, Lysis Buffer (5X)-Ultra, Dilution Buffer-Ultra, AlphaLISA® CaptSure™ Acceptor Beads 2mg/mL, Alpha Streptavidin Donor Beads 2mg/mL, Positive Control lysate.

STORAGE: Store all Buffers at 4°C.

Activation Buffer precipitates at 4°C. To re-dissolve, warm to 37°C and mix before use. Alternatively, store at room temperature with no loss in activity.

Store Positive Control Lysate at -20°C. Once reconstituted, the lysate will be stable for at least 1 month when stored at -20°C in single use aliquots.

Donor beads are used under low light conditions and should be stored in the box at 4°C. Vortex Acceptor and Donor Beads vigorously prior to taking samples for use.

KIT STABILITY: Stable for at least 9 months from the manufacturing date when stored in its original packaging and recommended storage conditions. This product will have a minimum of 3 months expiry from date of delivery. See kit box label for expiry date.

CONTROL LYSATES

EXPECTED SIGNAL: Reconstituted control lysate diluted to 25.00 % should give a signal window over buffer alone of at least 8 - fold.

The control lysate signal window can be influenced by lysate solubility, differing lab conditions, reader settings and reader performance.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

For more information regarding related SureFire Ultra products and protocols refer to PerkinElmer web site: <http://www.perkinelmer.com>

This product is not for resale or distribution except by authorized distributors.

LIMITED WARRANTY: TGR BioSciences warrants that, at the time of shipment, the products sold by it are free from defects in material and workmanship and conform to specifications which accompany the product. TGR BioSciences makes no other warranty, express or implied with respect to the products, including any warranty of merchantability or fitness for any particular purpose, or non-infringement of third party rights. Notification of any alleged breach of warranty must be made in writing within 60 days of receipt. No claim shall be honoured if the customer fails to notify TGR BioSciences within the period specified. The sole and exclusive remedy of the customer for any liability of TGR BioSciences of any kind including liability based upon warranty (express or implied whether contained herein or elsewhere), strict liability, contract or otherwise is limited to the replacement of the goods or the refunds of the invoice price of goods. TGR BioSciences shall not in any case be liable for special, incidental or consequential damages of any kind.

Worldwide Headquarters:
PerkinElmer, Inc.

940 Winter Street, Waltham, MA 02451 USA
Phone: (800) 762-4000 or (+1) 203-925-4602
Or locate your local sales office at www.perkinelmer.com

Please visit www.perkinelmer.com/Tools/COASearch
for Certificates of Analysis (and Technical Data)

