



## Certificate of Analysis

PerkinElmer  
Health Sciences B.V.  
Rigaweg 22  
P.O. Box 5205  
9700 GE Groningen  
The Netherlands

Phone: +31-50-544 59 00  
Fax: +31-50-544 59 50  
www.perkinelmer.com

**Product:** OptiPhase HiSafe 3

**Batch #:** 152-19311

**Exp. Date:** February 1, 2021

-----  
The product has been analyzed and is certified to have met the specifications.

Test	Temperature	Results
Appearance		Clear blue fluorescent liquid
Quench parameter		562
Background		18 cpm
Water Holding Capacity	20 °C	≥ 12.0 ml clear
1.0 M NaCl Holding Capacity	20 °C	≥ 8.0 ml clear
0.15 M NaCl Holding Capacity	20 °C	≥ 11.0 ml clear

Date: July 29, 2019

*S. Beukema*  
QA Supervisor  
sandra.beukema@perkinelmer.com  
Tel. + 31(0)50-5445925