

Anti-Rabbit Immunoglobulin G (rabIgG) *ULight* LANCE *Ultra* Toolbox

Product number: TRF502D / TRF502M / TRF502R

Lot number: 2726203

Caution: For Laboratory Use. A research product for research purposes only.

Material Provided

Format: TRF502D 0.2 nmole (1 000 assay points*)
TRF502M 2 nmoles (10 000 assay points*)
TRF502R 20 nmoles (100 000 assay points*)

*Assuming 0.2 pmol/ assay point

Volume: 400 µL (TRF502D), 4 x 1 mL (TRF502M) or 4 x 10 mL (TRF502R)

Manufacturing Date: May 27, 2020

Product Information

Application: *ULight* has been conjugated to anti-rabbit IgG antibody. This antibody recognizes rabbit immunoglobulins without considerations of isoforms or chain type. This toolbox can be used to either detect rabbit immunoglobulins in samples, or may be used as a secondary antibody to rabbit antibodies in TR-FRET assays, as rabbit are the most commonly encountered polyclonal antibodies in detection assays..

Storage: Store kit in the dark at +4°C.

Stability: This kit is stable for at least 6 months from the manufacturing date when stored in its original packaging and the recommended storage conditions

Quality Control

The QC release specifications are based on spectrophotometric analysis of the labeled antibody. We certify that results meet our quality release criteria.

Labeling Ratio: 4.0 (ULight/Proteins)

Concentration: 0.5 µM (80 µg/mL)

Description of the LANCE *Ultra* Assay

LANCE® and LANCE® (Lanthanide chelate excite) *Ultra* are our TR-FRET (time-resolved fluorescence resonance energy transfer), homogeneous (no wash) technologies. One antibody of interest is labeled with a donor fluorophore (a LANCE Europium chelate) and the second molecule is labeled with an acceptor fluorophore [*ULight*™ dye]. Upon excitation at 320 or 340 nm, energy can be transferred from the donor Europium chelate to the acceptor fluorophore if sufficiently close for FRET (~10 nm). This results in the emission of light at 665 nm.

Recommended Assay Conditions

Sodium azide should **not** be added to the stock reagents. High concentrations of sodium azide (> 0.001 % final in the assay) might decrease the signal.

Specific additional required reagents and materials:

The following materials are recommended:


Item	Suggested source	Catalog #
TopSeal™-A Plus Adhesive Sealing Film	PerkinElmer Inc.	6050185
Optiplate 96 or 384-well plate	PerkinElmer Inc.	6005290 (96) 6007299 (384)
VICTOR X, ViewLux, EnVision or EnSpire Multilabel Plate Reader equipped with TR-FRET option	PerkinElmer Inc.	-

Example: Anti-rabbit IgG *ULight* LANCE *Ultra* Assay


Reagents:

1. Prepare 1x Buffer: Add 2 mL of 5X PBS + 0.1% Tween 20 to 8 mL H₂O.
2. Prepare Biotin-rabbit IgG probe standard dilutions: Dilute Biotin-rabbit IgG Probe to 1 μM with 1X PBS + 0.1% Tween 20
3. Dilute Eu-W1024 Streptavidin (1 mg) to 500 nM with 1X TSA buffer (50 mM Tris-HCl – 150 mM NaCl – 0.05% sodium azide) pH 7.4

Protocol:



Typical Data



Recommendations

- The volume indicated on each tube is guaranteed for single pipetting. Multiple pipetting of the reagents may reduce the theoretical amount left in the tube.
- Centrifuge all tubes (including lyophilized analyte) before use to improve recovery of content (2000g, 10-15 sec).
- Re-suspend all reagents by vortexing before use.
- Use Milli-Q® grade H₂O (18 MΩ•cm) to dilute Buffer.
- When diluting the standard or samples, change tips between each standard or sample dilution. When loading reagents in the assay microplate, change tips between each standard or sample addition and after each set of reagents.
- When reagents are added to the microplate, make sure the liquids are at the bottom of the well.
- Small volumes may be prone to evaporation. It is recommended to cover microplates with TopSeal-A Adhesive Sealing Films to reduce evaporation during incubation. LANCE *Ultra* TR-FRET assays cannot be read with the TopSeal-A Film attached. Please remove before reading.
- LANCE signal is detected using a VICTOR X, ViewLux, EnVision or EnSpire Multilabel Reader equipped with the TR-FRET. Use an excitation wavelength of 320 or 340 nm to excite the LANCE Europium chelate. We recommend you read this assay in dual emission mode, detecting both the emission from the Europium donor fluorophore at 615 nm, and the acceptor fluorophore (at 665 nm for *ULight* dye). The raw FRET signal at 665 nm can be used to process your data.
- Signal will vary with temperature and incubation time. For consistent results, identical incubation times and temperature should be used for each plate.

Troubleshooting Guide

You will find detailed recommendations for common situations you might encounter with your LANCE *Ultra* Assay kit at:

<http://www.perkinelmer.com/Resources/TechnicalResources/ApplicationSupportKnowledgebase/LANCE/lance.xhtml>

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

This product is not for resale or distribution except by authorized distributors.

LIMITED WARRANTY: PerkinElmer BioSignal Inc. warrants that, at the time of shipment, the products sold by it are free from defects in material and workmanship and conform to specifications which accompany the product. PerkinElmer BioSignal Inc. makes no other warranty, express or implied with respect to the products, including any warranty of merchantability or fitness for any particular purpose. Notification of any breach of warranty must be made within 60 days of receipt unless otherwise provided in writing by PerkinElmer BioSignal Inc. No claim shall be honored if the customer fails to notify PerkinElmer BioSignal Inc. within the period specified. The sole and exclusive remedy of the customer for any liability of PerkinElmer BioSignal Inc. of any kind including liability based upon warranty (express or implied whether contained herein or elsewhere), strict liability contract or otherwise is limited to the replacement of the goods or the refunds of the invoice price of goods. PerkinElmer BioSignal Inc. shall not in any case be liable for special, incidental or consequential damages of any kind.

PerkinElmer, Inc.
940 Winter Street
Waltham, MA 02451 USA
P: (800) 762-4000 or
(+1) 203-925-4602
www.perkinelmer.com

For a complete listing of our global offices, visit www.perkinelmer.com/ContactUs

Copyright© 2012, PerkinElmer, Inc. All rights reserved. PerkinElmer® is a registered trademark of PerkinElmer, Inc. All other trademarks are the property of their respective owners.